

DATA SHEET-V4
REMBRANDT® LSI 3q26.2/CEP3-
AMPLIFICATION FISH DETECTION

RUO

Pan
Path

Ref	C814K.2030.05		5 T
	C814K.2030.10		10 T

mixture (formamide, dextran sulphate and SSC) and are ready to use solutions.

Intended use


- I. The REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay is for research use only and is intended for the 3q26.2 locus, compared to the centromeric region of chromosome 3 by means of *in situ* hybridization.
- II. The REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay is intended for the detection of the 3q26.2 locus, compared to the centromeric region of chromosome 3 in fixed cells. A clinical diagnosis should not be established based on the performance of this test.
- III. The intended users are qualified laboratory employees in cytology and/or pathology. The product is intended for professional use.

Clinical relevance

The TERC (telomerase reverse transcriptase) gene, also described as the human telomerase RNA component, is located in the genome at 3q26.2 and counts one exon. Telomerase is a ribonucleoprotein polymerase that preserves telomeric ends by addition of the telomere repeat 'TTAGGG' (TERC telomerase RNA component [Homo sapiens (human)], 2022) (NIH, n.d.). Telomerase is an enzyme consisting of two important components, the TERC gene encodes for the hTR component and the TERT gene encodes for the hTERT component. Mutations in the TERC gene lead to telomerase dysfunction resulting in compromised maintenance of telomeres and therefore telomeric degradation. TERC gene mutations are known to be present in for example autosomal dominant dyskeratosis congenita, idiopathic pulmonary fibrosis, and aplastic anemia (Shallis et al., 2018; Vulliamy et al., 2001).

Probe specification

The REMBRANDT® LSI 3q26.2/CEP3 probe mix consists of a 514 kb probe detecting the 3q26.2 locus and a 0.68 kb centromeric probe detecting the centromeric region of chromosome 3. The centromeric region is detected by green fluorescence (AF488) and the locus is detected by orange fluorescent detection (AF555). The REMBRANDT® LSI 3q26.2 probe is able to completely cover the TERC gene with flanking sequences on the 5' and 3' of the gene for signal enhancement. The REMBRANDT® LSI 3q26.2/CEP3 probes are pre-mixed in a hybridization



Test principle

In a fluorescent *in situ* hybridization assay, a double stranded DNA probe labelled with a fluorochrome is used. The labelled DNA probe is diluted in a hybridization mixture. The hybridization mixture containing the DNA probe is added to the specimen. A co-denaturation will ensure that the genomic DNA and the probe DNA become single stranded. After denaturation, the probe DNA is able to hybridize to its complementary target sequences in the cells. In the REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay, the fluorochrome is attached to the probe and the signals can be visualized directly by fluorescent microscopy after hybridization.

Reagents provided

Product name	Product number	Amount
Labelled probes (depending on size choice)		
REMBRANDT® LSI 3q26.2/CEP3 FISH probe mix	C814P.2030.05 or C814P.2030.10	5 T 10 T
REMBRANDT® Pepsin powder	R011R.0000	1 g
REMBRANDT® Pepsin diluent	R018R.0000	15 ml
REMBRANDT® PanWash 4, 25X SSC	R025R.0000	4x 15 ml
REMBRANDT® Fluorescent Mounting medium	Z000R.0050	1 ml

Assay procedure

REMBRANDT® LSI 3q26.2/CEP3 FISH detection assay procedure for cytological specimen.

- I. Incubate slides in pre-heated proteolytic work solution (prepare according to section 1.9 of Manual-FISH) (R011R.000 + R018R.000) at 37 °C (100 µg/ml) for 15 minutes followed by a brief rinsing in 0.01M HCl (1x 2 minutes) and subsequent rinses in PBS (2x 1 minute)
- II. Flush wash slides in deionised water, followed by dehydration in graded ethanol series (ethanol 70%, 96%, 96%, 100%, 100%) 1 minute each and air-dry slides for 15 minutes

Do not treat more than 5 slides at the same time, because the temperature of the pre-heated solutions may drop dramatically, thus causing incomplete pre-treatment. Additionally, allow the slides to air-dry as recommended; otherwise sections will be lost.

- III. Homogenize probe solution (C814P.2030.YY) a spin briefly. Apply 15 µl of probe solution to each specimen. Cover all specimens with a cover slip (avoid air bubbles). Denature slides on a 80 °C hotplate or other heating device for 3 minutes

Work in a pre-set order to ensure that all slides have been incubated at 80 °C for the exact same time. Do not denature more than 5 slides at the same time, the temperature of the heating device may drop dramatically, thus causing incomplete denaturation.

- IV. Transfer the slides into a moist environment and incubate for 16 hours at 37 °C.
- V. Remove coverslips by soaking the slides in PBS at room temperature
- VI. Incubate the slides in diluted, pre-heated PanWash 4 (R025R.000) (prepare according to section 1.9 of Manual-FISH)

Do not incubate more than 5 slides at the same time in PanWash 4, the temperature of the PanWash 4 may drop dramatically, causing wrong stringency conditions.

- VII. Incubate the slides in PBS at room temperature for 1 minute
- VIII. Dehydrate the slides in graded ethanol series (70%, 96%, 96%, 100%, 100%) 1 minute each and air-dry the slides for 15 minutes (in the dark)


Mount the slides by applying mounting medium (Z000R.0050) and coverslip

Interpretation of results

Hybridization of the REMBRANDT® LSI 3q26.2/CEP3 probes is viewed using a fluorescence microscope equipped with appropriate excitation and emission filters for orange detection: λ_{exc} 555 nm, λ_{em} 572 nm. Allowing visualization of orange fluorescent signal concentrated at the p11.2 locus of chromosome 3 in combination with green fluorescent signals representing the centromeric region of chromosome 3.

The assay procedure enables visual enumeration of the 3q26.2 locus compared to the centromeric region of chromosome 3 within the nuclei. Enumerate the fluorescent signals in the interphase nucleus using a 40X or 63X magnification. Objectives with higher magnification (eg, 63X or 100X) should be used to verify or resolve questions about split or diffused signals. Enumerate at least 100 nuclei per slide for accurate analysis.

- Two signals in close proximity and approximately the same sizes but not connected by a visible link are counted as 2 signals.
- Count a diffuse signal as 1 signal if diffusion of the signal is contiguous and within an acceptable boundary.
- Two small signals connected by a visible link are counted as 1 signal.
- Enumerate the number of nuclei with 0, 1, 2, 3, 4, or >4 signals. Count nuclei with 0 signals only if there are other nuclei with at least 1 signal present in the field of view. If the accuracy of the enumeration is in doubt, repeat the enumeration in another area of the slide.
- Do not enumerate nuclei with uncertain signals (Arsham et al., 2017)



Performance characteristics

Analytical Sensitivity and Specificity

The REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay was analytically validated for TERC amplification-FISH detection assay. The results of the direct fluorescent assay are shown. However, for the REMBRANDT® LSI 3q26.2/CEP3 amplification FISH

detection assay, the detection system may influence the performance characteristics and REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay in combination with different detection systems should be evaluated carefully by the end-user.

Analytical sensitivity

The analytical sensitivity was determined in three levels. The normal cut-off percentage was determined based on assessment of 200 individual nuclei in two samples of healthy interphase lymphocytes from peripheral blood. The beta inversion was used to determine the percentage of normal-cut off. For the noise-to-signal percentage, the noise and signal values were determined for 100 signals in interphase lymphocytes from peripheral blood of two independent samples. The differences between noise and signal were evaluated using a 95% confidence interval. For the hybridization efficiency, 200 individual nuclei were assessed for the presence of FISH signals.

Performance characteristic	Outcome
Normal cut-off percentage	10%
Noise-to-signal cut-off percentage	19%
Hybridization efficiency	99%

Analytical specificity

The analytical specificity was determined in two levels. The theoretical specificity was determined by sequencing analysis of the probe DNA and mapping on Hg38. The practical specificity was determined by assessing the hybridization pattern in metaphase chromosomes of interphase lymphocytes from peripheral blood.

Performance characteristic	Outcome
Theoretical specificity	Mapped on chromosome 3, locus q26.2 and the centromeric region p11.1-q11.1
Practical specificity	100%

Limitations of Procedure

i) The REMBRANDT® LSI 3q26.2/CEP3 amplification FISH detection assay is solely applicable for the detection of the 3q26.2 locus and the centromeric region of chromosome 3, which may be present in cell preparations (cytological specimen i.e. interphase lymphocytes from peripheral blood samples).

ii) Either human tissue sections or human cytological preparations may be used. Samples must be fixed in buffered formalin or alcohol. In tissue sections are required, the sections should be prepared in a 4 µm thickness. Furthermore, the tissues should be glued to the glass slides with a bio-adhesive (e.g. organ silane), dried at room temperature, subsequently dried at 37 °C overnight and lastly completely deparaffinized in xylene and alcohol series and air dried.

iii) Cytological specimen should be prepared as required by the user, fixed with cytological fixation agent, rinsed in distilled water prior to the ISH procedure and air dried.

iv) Many factors can influence the performance of the ISH procedure. Failure in detection can be due to i.e. improper sampling, handling, the time lapse between tissue removal and fixation, the size of the tissue specimen in the fixation medium, the fixation time, processing fixed tissue, the thickness of the section, the bio-adhesive on the slide, deparaffinisation procedure, incubation times, proteolytic pre-treatment, detection reagents, incubation temperatures and interpretation of results.

v) The performance of the ISH procedure is also affected by the sensitivity of the method and the presence of the locus 3q26.2 and the centromeric region of chromosome 3. In case the limit of the sensitivity is reached a false negative reaction may be the result.

viii) Therapeutic considerations based on the result of this test alone should not be taken. Results should be verified by other traditional diagnostic methods such as but not limited to clinical history, symptoms, as well as morphological data.

ix) The medical profession should be aware of risks and factors influencing the fluorescent signal intensity while interpreting the test result. Microscopy settings might influence the signal intensity and/or interpretation.

x) Laboratory personnel performing the test should be trained and knowledgeable to be able to interpret the test results.

Storage and handling

Store kit and its contents at 2-8°C. Store the dissolved and aliquoted reagents at recommended temperatures. When used and stored as indicated, the kit is stable until the expiry date printed on the box.

Product	Product number	Storage conditions
REMBRANDT® LSI 3q26.2/CEP3 FISH probe mix	C814P.2030.YY	2-8 °C
REMBRANDT® Pepsin powder	R011R.0000	Powder: 2-25 °C, ambient temperature
REMBRANDT® Pepsin diluent	R018R.0000	Dissolved: -20 °C Concentrated solution and diluted: 2-25 °C, ambient temperature

REMBRANDT® R025R.0000 Concentrated solution and diluted: 2-25 °C, ambient temperature 2-8 °C

REMBRANDT® Z000R.0050 Fluorescent mounting medium



Hazard statements

H315 - Causes skin irritation
H319 - Causes serious eye irritation
H351 - Suspected of causing cancer
H360D - May damage the unborn child
H373 - May cause damage to organs through prolonged or repeated exposure

Precautionary Statements

P202 - Do not handle until all safety precautions have been read and understood
P280 - Wear protective gloves/protective clothing/eye protection/face protection
P302 + P352 - IF ON SKIN: Wash with plenty of water and soap
P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing
P308 + P313 - IF exposed or concerned: Get medical advice/attention
P362 + P364 - Take off contaminated clothing and wash it before reuse
P405 - Store locked up

Additional information

Product in combination with other devices

The REMBRANDT® DNA probes are intended for stand-alone usage. The assay is intended to be used in combination with standard cytological preparation methods, hot plate(s), stove(s), incubation device(s), water bath(s), temperature and incubation time control(s), proteolytic-, detection- and other reagents (not supplied with this reagent) and a microscope. The combination has been tested and validated. Since the standard cytological preparation methods, hot plate(s), stove(s), incubation device(s), water bath(s), temperature controls, incubation time control(s) and other not supplied reagents such as but not limited to proteolytic reagents, detection reagents and a microscope is not combined with the device as a product, conformity with the essential requirements is not applicable. Assay validation criteria are mentioned in 'Interpretation of the Results' and are also depending on clinical state of the sample, which may influence the validation criteria.

For additional information regarding the REMBRANDT® assays, a manual is included which specifies the following subjects:

- Controls
- Materials required but not included
- Storage and shelf-life
- Performance precautions
- Preparations of reagents
- Specimen collection
- Quality control
- Trouble shooting guide

Technical assistance

For technical assistance regarding the products performance, please contact info@panpath.nl or call +31 495499090. Visit our website for reprints of datasheets or additional documentation. www.panpath.nl

Literature list



- Arsham, M. S., Barch, M. J., & Lawce, H. J. (2017). *The AGT Cytogenetics Laboratory Manual The AGT Cytogenetics Laboratory Manual Edited by* (Vol. 4).
- MedlinePlus, National Library of Medicine.* (2020, 07 1). Opgehaald van TERC gene, telomerase RNA component: <https://medlineplus.gov/genetics/gene/terc/>
- Rory M Shallis, R. A. (2018). Aplastic anemia: Etiology, molecular pathogenesis, and emerging concepts. *European Journal of Haematology*, 711-720.
- T Vulliamy, A. M. (2001). The RNA component of telomerase is mutated in autosomal dominant dyskeratosis congenita. *Nature*, 432-436.
- TERC telomerase RNA component [Homo sapiens (human)]*. (2022, 02 13). Opgehaald van National Center for Biotechnology Information (NCBI)[Internet]. Bethesda (MD): National Library of Medicine (US), National Center for Biotechnology Information: <https://www.ncbi.nlm.nih.gov/gene?Db=gene&Cmd=DetailsSearch&Term=7012>

Disclaimer: This document is valid until the product expiry on the kit label

DATA SHEET-V2
REMBRANDT® LSI 3q26.2/CEP3
AMPLIFICATION ISH DETECTION
RESEARCH USE ONLY (RUO)

RUO

PanPath

Ref **C814K.0199.05**  5 T
C814K.0199.10  10 T

Intended use

- I. The REMBRANDT® LSI 3q26.2/CEP3 amplification ISH detection assay is intended for the detection of the 3q26.2 locus, compared to the centromeric region of chromosome 3 by means of *in situ* hybridization.
- II. The REMBRANDT® LSI 3q26.2/CEP3 amplification ISH detection assay is intended for the detection of the 3q26.2 locus, compared to the centromeric region of chromosome 3 in fixed cells. A clinical diagnosis should not be established based on the performance of this test.
- III. The intended users are qualified laboratory employees in cytology and/or pathology. The product is intended for professional use.


Clinical relevance

The TERC (telomerase reverse transcriptase) gene, also described as the human telomerase RNA component, is located in the genome at 3q26.2 and counts one exon. Telomerase is a ribonucleoprotein polymerase that preserves telomeric ends by addition of the telomere repeat 'TTAGGG' (TERC telomerase RNA component [Homo sapiens (human)], 2022) (NIH, n.d.). Telomerase is an enzyme consisting of two important components, the TERC gene encodes for the hTR component and the TERT gene encodes for the hTERT component. Mutations in the TERC gene lead to telomerase dysfunction resulting in compromised maintenance of telomeres and therefore telomeric degradation. TERC gene mutations are known to be present in for example autosomal dominant dyskeratosis congenita, idiopathic pulmonary fibrosis, and aplastic anemia (Shallis et al., 2018; Vulliamy et al., 2001).

Probe specification

The LSI 3q26.2/CEP3 probes set consists of a 514 kb probe detecting the 3q26.2 locus and a 0.68 kb centromeric probe detecting the centromeric region of chromosome 3. The centromeric region is conjugated to biotin and the locus is conjugated to digoxigenin. The LSI 3q26.2 probe is able to completely cover the TERC gene with flanking sequences on the 5' and 3' of the gene for signal enhancement. The EGFR amplification probes are



pre-mixed in a hybridization mixture (formamide, dextran sulphate and SSC) and are ready to use solutions.



Test principle

In an *in situ* hybridization assay, a double stranded DNA probe labelled with a hapten is used. The labelled DNA probe is diluted in a hybridization mixture. The hybridization mixture containing the DNA probe is added to the specimen. A co-denaturation will ensure that the genomic DNA and the probe DNA become single stranded. After denaturation, the probe DNA is able to hybridize to its complementary target sequences in the cells. In the REMBRANDT® LSI 3q26.2/CEP3-ISH assay, the haptens are attached to the probe and the signals can be visualized after detection by corresponding antibodies by fluorescent or brightfield microscopy.

Reagents provided

Product name	Product number	Amount
Labelled probe (depending on label and size choice)		
REMBRANDT® LSI 3q26.2/CEP3-ISH probe mix	C814P.0199.05 or C814P.0199.10	 5 T  10 T
REMBRANDT® Pepsin powder	R011R.0000	1 g
REMBRANDT® Pepsin diluent	R018R.0000	15 ml
REMBRANDT® Pre-treatment solution	R026R.0000	15 ml
REMBRANDT® PanWash 4, 25X SSC	R025R.0000	15 ml

Assay procedure

REMBRANDT® LSI 3q26.2/CEP3 amplification ISH detection assay procedure for cytological specimen and FFPE tissue sections.

- I. Specimen collection: for a detailed description of the specimen collection for cytological specimen or FFPE tissue sections see section 2.1 Specimen collection of the Manual ISH.
- II. For FFPE tissue sections, after dewaxing, place slides in jar with pre-treatment solution (R026R.0000) in microwave set at i.e. 900W and incubate up until boiling. Subsequently, reset microwave at 180W and incubate for 10 minutes, followed by cooling down for 20 minutes at room temperature. Flush wash slides in deionised water.
- III. Incubate slides in pre-heated proteolytic work solution (prepare according to section 1.9 and 1.10 of Manual ISH (R011R.000 + R018R.000) at 37 °C. Paraffin-embedded sections (1.25 mg/ml) or cytological specimen (100 µg/ml) for 15 minutes.
- IV. Flush wash slides in deionised water, followed by dehydration in graded ethanol series (ethanol 70%, 96%, 96%, 100%, 100%) 1 minute each and air-dry slides for 15 minutes.

Do not treat more than 5 slides at the same time, because the temperature of the pre-heated solutions may drop dramatically, thus causing incomplete pre-treatment. Additionally, allow the slides to air-dry as recommended; otherwise sections will be lost.

- V. Homogenize probe solution (C814P.0199.YY) and spin briefly. Apply 10-15 µl of probe solution to each specimen. Cover all specimens with a cover slip (avoid air bubbles). Denature slides on a 80 °C hotplate or other heating device, 3 minutes for cytological specimen and 10 minutes for FFPE tissue sections.

Work in a pre-set order to ensure that all slides have been incubated at 80 °C for the exact same time. Do not denature more than 5 slides at the same time, the temperature of the heating device may drop dramatically, thus causing incomplete denaturation.

- VI. Transfer the slides into a moist and dark environment and incubate for 16 hours at 37 °C.
- VII. Remove coverslips by soaking the slides in PBS at room temperature.
- VIII. Incubate the slides in diluted, pre-heated PanWash 4 (R025R.0000) (prepare according

to section 1.9 of Manual-ISH). For cytological specimen and FFPE tissue sections, 2x 5 minutes in 2x SSC at 42 °C. For cytological specimen, subsequently incubate 2x 5 minutes in 0.1x SSC at 61 °C.

Do not incubate more than 5 slides at the same time in PanWash 4, the temperature of the PanWash 4 may drop dramatically, causing wrong stringency conditions.

- IX. Appropriate detection system should be evaluated by the end-user. Recommended detection systems are listed below

Digoxigenin detection	Biotin detection
R003R.0000 REMBRANDT® Sheep aDig-AP conjugate	R041R.0000 REMBRANDT® Goat aBio-AP Fab conjugate
R004R.0000 REMBRANDT® Sheep aDig-HRP conjugate	R042R.0000 REMBRANDT® Goat aBio-HRP Fab conjugate
AP detection	HRP detection
R008R.0000 REMBRANDT® NBT/BCIP substrate	R007R.0000 REMBRANDT® AEC substrate + R010R.0000 REMBRANDT® AEC buffer

Interpretation of results

Hybridization of the LSI 3q26.2/CEP3 probe is viewed using a fluorescence or bright field microscope equipped with appropriate excitation and emission filters. Allowing visualization of ISH signals concentrated at the 3q26.2 locus, and the centromeric region of chromosome 3. The enumeration of the locus 3q26.2 is conducted by microscopic examination of interphase nuclei.

The assay procedure enables visual enumeration of the 3q26.2 locus, compared to the centromeric region of chromosome 3 within the nuclei. Enumerate the fluorescent signals in the interphase nucleus using a 40X or 63X magnification. Objectives with higher magnification (eg. 63X or 100X) should be used to verify or resolve questions about split or diffused signals. Enumerate at least 100 nuclei per slide for accurate analysis. Enumerate the fluorescent signals in the interphase nucleus using a 40X or 63X magnification. Objectives with higher magnification (eg. 63X or 100X) should be used to verify or resolve questions about split or diffused signals. Enumerate at least 100 nuclei per slide for accurate analysis.


•Two signals in close proximity and approximately the same sizes but not connected by a visible link are counted as 2 signals.

•Count a diffuse signal as 1 signal if diffusion of the signal is contiguous and within an acceptable boundary.

•Two small signals connected by a visible link are counted as 1 signal.

•Enumerate the number of nuclei with 0, 1, 2, 3, 4, or >4 signals. Count nuclei with 0 signals only if there are other nuclei with at least 1 signal present in the field of view. If the accuracy of the enumeration is in doubt, repeat the enumeration in another area of the slide.

•Do not enumerate nuclei with uncertain signals



Performance characteristics

Analytical Sensitivity and Specificity

The LSI 3q26.2/CEP3-ISH detection assay was analytically validated for LSI 3q26.2/CEP3-FISH detection. The results of the direct fluorescent assay are shown. However, for the REMBRANDT® LSI 3q26.2/CEP3-ISH detection assay, the detection system may influence the performance characteristics and LSI 3q26.2/CEP3-ISH detection assay in combination with different detection systems should be evaluated carefully by the end-user.

Analytical sensitivity

The analytical sensitivity was determined in three levels. The normal cut-off percentage was determined based on assessment of 200 individual nuclei in two samples of healthy interphase lymphocytes from peripheral blood. The beta inversion was used to determined the percentage of normal-cut off. For the noise-to-signal percentage, the noise and signal values were determined for 100 signals in interphase lymphocytes from peripheral blood of two independent samples. The differences between noise and signal were evaluated using a 95% confidence interval. For the hybridization efficiency, 200 individual nuclei were assessed for the presence of ISH signals.

Performance characteristic	Outcome
Normal cut-off percentage	10%

Noise-to-signal cut-off percentage	19%
Hybridization efficiency	99.5%

Analytical specificity

The analytical specificity was determined in two levels. The theoretical specificity was determined by sequencing analysis of the probe DNA and mapping on Hg38. The practical specificity was determined by assessing the hybridization pattern in metaphase chromosomes of interphase lymphocytes from peripheral blood.

Performance characteristic	Outcome
Theoretical specificity	Mapped on chromosome 3, q26.2 and the centromeric region p11.1-q11.1
Practical specificity	100%

Limitations of Procedure

i) The REMBRANDT® LSI 3q26.2/CEP3 amplification ISH detection assay is solely applicable for the detection of the locus 3q26.2 and the centromeric region of the human chromosome 3, which may be present in cell preparations (cytological specimen i.e. interphase lymphocytes from peripheral blood samples).

ii) Either human tissue sections or human cytological preparations may be used. Samples must be fixed in buffered formalin or alcohol. In tissue sections are required, the sections should be prepared in a 4 µm thickness. Furthermore, the tissues should be glued to the glass slides with a bio-adhesive (e.g. organ silane), dried at room temperature, subsequently dried at 37 °C overnight and lastly completely deparaffinized in xylene and alcohol series and air dried.

iii) Cytological specimen should be prepared as required by the user, fixed with cytological fixation agent, rinsed in distilled water prior to the ISH procedure and air dried.

iv) Many factors can influence the performance of the ISH procedure. Failure in detection can be due to i.e. improper sampling, handling, the time lapse between tissue removal and fixation, the size of the tissue specimen in the fixation medium, the fixation time, processing fixed tissue, the thickness of the section, the bio-adhesive on the slide, deparaffinisation procedure, incubation times, proteolytic pre-treatment, detection reagents, incubation temperatures and interpretation of results.

v) The performance of the ISH procedure is also affected by the sensitivity of the method and the presence of the locus 3q26.2 and the centromeric region of chromosome 3. In case the limit of the sensitivity is reached a false negative reaction may be the result.

vi) The REMBRANDT® LSI 3q26.2/CEP3 amplification ISH detection assay results should not be relied on in case the sampling, sampling method, sample quality, sample preparation, reagents used, controls and procedure followed are not optimal or as described the working protocols.

vii) The medical profession should be aware of risks and factors influencing the fluorescent signal intensity while interpreting the test result. Microscopy settings might influence the signal intensity and/or interpretation.

viii) Laboratory personnel performing the test should be trained and knowledgeable to be able to interpret the test results.

Storage and handling

Store kit and its contents at 2-8°C. Store the dissolved and aliquoted reagents at recommended temperatures. When used and stored as indicated, the kit is stable until the expiry date printed on the box.

Product	Product number	Storage conditions
REMBRANDT® LSI 3q26.2/CEP3-ISH probe mix		2-8 °C
REMBRANDT® PanWash 4, 25X SSC	R025R.0000	Concentrated solution and diluted: 2-25°C, ambient temperature
REMBRANDT® Pepsin powder	R011R.0000	Powder: 2-25°C, ambient temperature Dissolved: -20°C
REMBRANDT® Pepsin diluent	R018R.0000	Concentrated solution and diluted: 2-25°C, ambient temperature
REMBRANDT® Pre-treatment buffer	R026R.0000	Concentrated solution and diluted: 2-25°C, ambient temperature



Hazard statements

H315 - Causes skin irritation
H319 - Causes serious eye irritation
H351 - Suspected of causing cancer
H360D - May damage the unborn child
H373 - May cause damage to organs through prolonged or repeated exposure

Precautionary Statements

P202 - Do not handle until all safety precautions have been read and understood
P280 - Wear protective gloves/protective clothing/eye protection/face protection
P302 + P352 - IF ON SKIN: Wash with plenty of water and soap
P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to

do. Continue rinsing
P308 + P313 - IF exposed or concerned: Get medical advice/attention
P362 + P364 - Take off contaminated clothing and wash it before reuse
P405 - Store locked up

Additional information

Product in combination with other devices

The REMBRANDT® DNA probes are intended for stand-alone usage. The assay is intended to be used in combination with standard cytological preparation methods, hot plate(s), stove(s), incubation device(s), water bath(s), temperature and incubation time control(s), proteolytic-, detection- and other reagents (not supplied with this reagent) and a microscope. The combination has been tested and validated. Since the standard formalin fixed, paraffin embedded tissue blocks, standard tissue freezing, tissue sectioning (microtome), standard cytological preparation methods, hot plate(s), stove(s), incubation device(s), water bath(s), temperature controls, incubation time control(s) and other not supplied reagents such as but not limited to proteolytic reagents, detection reagents and a microscope is not combined with the device as a product, conformity with the essential requirements is not applicable. Assay validation criteria are mentioned in 'Interpretation of the Results' and are also depending on the target load, which may influence the validation criteria.

For additional information regarding the REMBRANDT® assays, a manual is included which specifies the following subjects:

- Controls
- Materials required but not included
- Storage and shelf-life
- Performance precautions
- Preparations of reagents
- Specimen collection
- Quality control
- Trouble shooting guide

Technical assistance

For technical assistance regarding the products performance, please contact info@panpath.nl or call +31 495499090. Visit our website for reprints of datasheets or additional documentation. www.panpath.nl

Literature list

Arsham, M. S., Barch, M. J., & Lawce, H. J. (2017). *The AGT Cytogenetics Laboratory Manual The AGT Cytogenetics Laboratory Manual Edited by* (Vol. 4).
MedlinePlus, National Library of Medicine. (2020, 07 1).
Opgehaald van TERC gene, telomerase RNA component:
<https://medlineplus.gov/genetics/gene/terc/>

Rory M Shallis, R. A. (2018). Aplastic anemia: Etiology, molecular pathogenesis, and emerging concepts. *European Journal of Haematology*, 711-720.

T Vulliamy, A. M. (2001). The RNA component of telomerase is mutated in autosomal dominant dyskeratosis congenita. *Nature*, 432-436.

TERC telomerase RNA component [Homo sapiens (human)]. (2022, 02 13). Opgehaald van National Center for Biotechnology Information (NCBI)[Internet]. Bethesda (MD): National Library of Medicine (US), National Center for Biotechnology Information: <https://www.ncbi.nlm.nih.gov/gene?Db=gene&Cmd=DetailsSearch&Term=7012>